



CERTIFICATE OF ANALYSIS

Certified Reference Material

Assay: RNaseP

Description: RT-PCR Master Mix, Frozen, 40 reactions per tube.

Lot Number: 270614-03M

Period of Validity: Expires June 2016

Certified Property Values:

Test	Specifications	Results
Contamination	Master mix only replicates produce no amplification	PASS
Identification	Master mix detected assay specific positive control	PASS
Homogeneity	Master mix is sufficiently homogeneous	PASS

Certification Analytical Method:

Real-time Reverse Transcriptase PCR on the JBAIDS platform was used to validate the certification values of CRM. Homogeneity was established by evaluating a minimum of 80 samples of the reference material in its final sellable state by a minimum of two technicians on two separate days with two instruments if available. Protocols have been followed to insure the certified value represents the true value. NMRC cannot be held responsible for changes that happen during storage of the master mix at the customer's location.

Intended Use:

CRM is to be used for *in vitro* diagnostic use only under EUA for the presumptive detection of Ebola Zaire virus (detected in the West Africa outbreak in 2014) on specified instruments in individuals in affected areas with signs and symptoms of Ebola virus infection or who are at risk for exposure or may have been exposed to the Ebola Zaire virus (detected in the West Africa outbreak in 2014) in conjunction with epidemiological risk factors.

Instructions for Use:

Refer to instruction booklet for details regarding proper use and storage conditions.

Hazards: CRM is considered non-hazardous. MSDS is available upon request.

RNaseP RT-PCR Master Mix lot 270614-03M has been certified for use on 9 July 2014 by Joan Gebhardt, Head, PCR Production Facility.



Naval Medical Research Center PCR Production Facility
Biological Defense Research Directorate
8400 Research Plaza
Fort Detrick, MD 21702